



GXR-4 Copper-Mill Heads

Summary

Precise and rapid determination of total mercury in GXR-4 can be performed using Direct Mercury Analyzer. Such an instrument requires no sample wet chemistry or pre-treatment.

Once a weighed sample portion is introduced into the instrument, analysis is completed in six minutes. Direct analysis of mercury, using the integrated sequence of Thermal Decomposition, Catalyst Conversion, Amalgamation, and Atomic Absorption Spectrophotometer, is described in EPA 7473 and is validated for laboratory as well as field analysis.

Instrumentation

Direct Mercury Analyzer apparatus and supplies

Milestone DMA-80, 640-1640 terminal with DMA-80 software or DMA-80 PC software, metal boats.

Analytical balance, spatula, pipette, or appropriate mechanical pipette and volumetric flask(Class A), 50 or 100 ml.

Sample weight :

Up to 100 mg (max)

The sample has been mixed and then introduced into the metal boat.

Procedure

1. Place a boat on the balance plate, tare it and weigh the sample.
2. Introduce the boat into sample tray.
3. Run the DMA-80 program to completion.

DMA-80 program

N° step	Time	Temperature
1	00:01:00	200°C
2	00:02:00	650°C
3	00:01:00	650°C

Max start temperature: 200°C
Purge: 60 sec

Results

Results µg/kg	Statistical Data		
108.44-111.93	<i>avg:</i>	<i>sd:</i>	<i>rsd:</i>
107.18-104.42	107.76	2.46 µg/kg	2.28 %
106.81	µg/kg		

Conclusion

The DMA-80 Mercury Analyzer successfully processed GXR-4 sample. Total analysis time per sample was less than 6 minutes, including the time taken to weigh each sample into the boat.