



### Summary

Precise and rapid determination of total mercury in Crude oil can be performed using DMA-1 Direct Mercury Analyzer. Such an instrument requires no sample wet chemistry or pre-treatment.

Once a weighed sample portion is introduced into the instrument, analysis is completed in six minutes. Direct analysis of mercury, using the integrated sequence of Thermal Decomposition, Catalyst Conversion, Amalgamation, and Atomic Absorption Spectrophotometer, is described in EPA 7473 and is validated for laboratory as well as field analysis.

### Instrumentation

#### *Direct Mercury Analyzer apparatus and supplies*

Milestone DMA-1, 640-12 and 1640-16 terminal with DMA-1 software or DMA-1 PC software, quartz boats.

Analytical balance, spatula, pipette, or appropriate mechanical pipette and volumetric flask(Class A), 50 or 100 ml.

### Sample weight :

50 mg

Due to the high organic of crude oil sample, the largest sample size that should be analyzed by this instrument is 50mg

### Additive:

Sample Adsoquick (supplied by Milestone)

### Procedure

1. Place a boat on the balance plate, tare it and weigh 300 mg of Adsoquick.
2. Add the sample.
3. Introduce the boat into the System.
4. Run the DMA-1 program to completion.

### DMA-1 program

N° step	Time	Temperature
1	00:01:00	200°C
2	00:02:00	650°C
3	00:01:00	650°C
Max start temperature: 200°C		
Purge: 60 sec		

### Results

sample	mg/kg
1	2.7 ± 0.9
2	22 ± 7

### Conclusion

Taking into account small sample size and viscosity of the sample, to increase precision the weight and not the volume should be reported. Hg concentration also should be calculated based on the sample weight. This is an ideal application for operation of the DMA-1 Mercury Analyzer in the field. With proper consideration for electrical power, and a firm platform on which to place the instrument and analytical balance, the DMA-1 Mercury Analyzer could easily be used on-site to evaluate the crude oil.